



**Dresser™ Meters & Instruments – List of Acceptable Dry Gases
to be run in
Dresser™ Rotary Gas Meters**

METERS OF STANDARD CONSTRUCTION

- Air
- Argon
- Butane
- Carbon Dioxide (dry)
- Carbon Monoxide (dry)
- Ethane
- Ethylene (C₂H₄)
- Helium
- Methane (Natural Gas)
- Neon
- Nitrogen
- Propane

METERS OF SPECIAL CONSTRUCTION

- Sewage or Digester Gas (Requires SSM Version – Special Service Meter)

NON-APPROVED DRY GASES

The following gases are not approved for use in any Dresser meter.

- Hydrogen
- Oxygen

Note: This list is subject to change without notice.



Distributed by:
IMAC Systems, Inc.
90 Main Street, Tullytown, PA. 19007
t: 215 946-2200 f: 215 943-2984
e: sales@imacsystems.com
www.imacsystems.com